

PRODUCT ALERT

PRODUCT ALERT

PRODUCT ALERT

## URGENT: DEVICE RECALL

One lot of **ThermaCare® HeatWraps Menstrual** was recalled on September 24, 2010 due to potential for a component leak which could result in 2<sup>nd</sup> degree skin burns. If you have purchased this product and have it in your home, please check to see if it has Lot Number E06831. If so, throw the product away without opening. This number may be followed by an S or N.

Find Lot  
Number  
Here →



Find Lot  
Number  
Here →



For more information, consumers can visit [www.thermacare.com](http://www.thermacare.com) or call the Product Information Line at 1-800-323-3383, Monday through Friday, 9am to 5pm EST for product replacement.

This notice for posting up until February 2011